SELEC
Creating Best Value

CE

EC Declaration of Conformity

Selec Controls Pvt. Ltd.
EL-27/1, Electronic zone, TTC Industrial Area,
Mahape, Navi Mumbai, India-400710.

Declares that the DOC is issued under our sole responsibility and belong to the following Product / Product Series

Product Descriptions	SMPS
Product Name / Series	OPS2x4-60-XXX-VCYY-Z-CU XXX- Represents the typical output voltage in VDC V- Represents the input voltage range (Blank for 100-240 VAC and 3 for 100-277 VAC) C- Represents the connector type (Any alphabet indicating type/ make of output connector) YY- Represents the Minor output voltage variation (Blank – Standard product or Any alphanumeric combination indicating minor minor output voltage variation in specification) Z- Represents classification of the device (1 for Class I (with earth) and 2 for Class II (without earth)

Electromagnetic Compatibility Directive (EMC) : 2014/30/EU		
Conducted Emission	CISPR 32:2015+AMD1:2019	CLASS B
Radiated Emission	CISPR 32:2015+AMD1:2019	CLASS A
ESD Immunity	IEC 61000-4-2:2008	Level 4, Criteria A Level 3, Criteria A
Radiated Field Immunity	IEC 61000-4-3:2020	Level 3, Criteria B
Electrical Fast Transient Immunity	IEC 61000-4-4:2012	Level 3, Criteria A
Surge Immunity	IEC 61000-4-5:2014 + AMD1: 2017	Level 3, Criteria A
Conducted Immunity	IEC 61000-4-6:2023	Level 3, Criteria B
Magnetic Field Immunity	IEC 61000-4-8:2009	Level 4, Criteria A
Voltage dips, interruptions	IEC 61000-4-11:2020 + COR2: 2022	Criteria A & B
Low Voltage Directive (LVD) : 2014/35/EU		
RoHS Directive : 2015/863/EU		

The following harmonized standards and technical specifications have been applied

EN IEC 60601-1:2005, AMD1:2012

EN IEC 62368-1:2023

Signed for and on behalf of: Selec Controls Pvt. Ltd.

Navi Mumbai
08/11/2023

Place and date of issue

SELEC CONTROLS PVT. LTD. NAVI MUMBAI 400710

Mr. Sanjay P., Design H.O.D.
(Name, function)